



Enzyme Research Laboratories

1801 Commerce Drive
South Bend, IN U.S.A. 46628
(574) 288-2268 * Fax (574) 288-2272
Website: www.enzymeresearch.com
Email: info@enzymeresearch.com

REPRESENTATIVE DATA SHEET

Matched-Pair Antibody Set for ELISA of human Factor XI antigen (F.XI)

Sufficient reagent for **5 x 96 well** plates

Product #: FXI-EIA
Lot #: SAMPLE
Expiry Date: SAMPLE

Store at -10 to -20°C

For Research Use Only
Not for use in diagnostic procedures.

Description of Factor XI (F.XI)

Factor XI (F.XI, plasma thromboplastin antecedent) is a coagulation protein produced in the liver that circulates in plasma at approximately 5 µg/ml (30 nM). The mass of F.XI is 160 kDa as determined by SDS-PAGE under non-reducing conditions and 80 kDa upon reduction. F.XI consists of two identical 80 kDa subunits linked by disulphide bonds. Each subunit consists of a tandem repeat of four apple domains followed by a serine protease catalytic domain. Cleavage of F.XI by activated factor XII or thrombin converts each subunit into a two-chain form and generates two active sites per F.XIa molecule. The mass of F.XIa is 160 kDa unreduced, but upon reduction F.XIa migrates as a heavy chain of 50 kDa and a light chain of 30 kDa. The catalytic site of F.XIa resides in the light chain. In plasma, F.XI or F.XIa circulates in non-covalent 1:1 complex with high molecular weight kininogen, which acts as a cofactor in the activation of F.XI by activated factor XII. The activity of F.XIa is regulated by platelets and by several proteinase inhibitors including, in order of decreasing importance, C1-inhibitor, α_2 antiplasmin, α_1 antitrypsin and antithrombin. Heparin has relatively little effect on the rate of inhibition of F.XIa by antithrombin. The only known natural substrate for activated F.XI (F.XIa) is factor IX (Christmas factor) and the only cofactor required for this reaction is ionized calcium¹⁻³.

Principle of Sandwich-style ELISA

Affinity-purified antibody to F.XI is coated onto the wells of a microtitre plate. Any remaining binding sites on the plastic wells are blocked with an excess of bovine serum albumin. The plates are washed and plasma or other fluids containing F.XI are applied. The coated antibody will capture the F.XI in the sample. After washing the plate to remove unbound material, a peroxidase conjugated second antibody to F.XI is added to the plate to bind to the captured F.XI. After washing the plate to remove unbound conjugated antibody, the peroxidase activity is expressed by incubation with o-phenylenediamine (OPD). After a fixed development time the reaction is quenched with the addition of H₂SO₄ and the colour produced is quantified using a microplate reader. The colour generated is proportional to the concentration of F.XI present in the sample.

Supplied Materials:

- 1. Capture Antibody (FXI-EIA-C):** One yellow-capped vial containing 0.5 ml of polyclonal affinity purified anti-F.XI antibody for coating plates.
- 2. Detecting Antibody (FXI-EIA-D):** One red-capped vial containing 0.5 ml of peroxidase conjugated polyclonal anti-F.XI antibody for detection of captured F.XI.

Note: Antibodies are supplied in a 50% (v/v) glycerol solution for storage at -10 to -20°C. Keep vials tightly capped. Do not store in frost-free freezers.

Materials Required but not Provided:

- 1. Coating Buffer:** 50 mM Carbonate
1.59g of Na₂CO₃ and 2.93g of NaHCO₃ up to 1 litre. Adjust pH to 9.6. Store at 2-8°C up to 1 month.
- 2. PBS:** (base for wash buffer and blocking buffer)
8.0g NaCl, 1.15g Na₂HPO₄, 0.2g KH₂PO₄ and 0.2g KCl, up to 1 litre. Adjust pH to 7.4, if necessary. Store up to 1 month at 2-8°C, discard if there is evidence of microbial growth.
- 3. Wash Buffer:** PBS-Tween (0.1%,v/v)
To 1 litre of PBS add 1.0 ml of Tween-20.
Check that the pH is 7.4. Store at 2-8°C up to 1 week.
- 4. Blocking Buffer:** PBS-BSA (1%, w/v)
Dissolve 2.5 g of Bovine Serum Albumin (Sigma-RIA grade) in 200 ml of PBS. Adjust pH to 7.4, if required, then make up to 250 ml with PBS. Aliquot and store frozen at -20°C.
- 5. Sample Diluent:** HBS-BSA-T20
5.95g HEPES (free acid), 1.46 g NaCl, 2.5 g Bovine Serum Albumin (Sigma, RIA grade) dissolved in 200 ml H₂O. Add 0.25 ml of Tween-20, check and adjust pH to 7.2 with NaOH, then make up to a final volume of 250 ml with H₂O. Aliquot and store frozen at -20°C.
- 6. Substrate Buffer:** Citrate-Phosphate buffer pH 5.0
2.6g Citric acid and 6.9g Na₂HPO₄ up to a final volume of 500 ml with purified H₂O. Store at 2-8°C up to 1 month.
- 7. OPD Substrate:** (o-Phenylenediamine.2HCl) **TOXIC!**
(5mg tablets: Sigma # P-6912). Make up immediately before use. Dissolve 5mg OPD in 12 ml substrate buffer then add 12 µl 30% H₂O₂. Do not store.

8. Stopping Solution: 2.5 M H₂SO₄

Caution: VERY CORROSIVE! GENERATES HEAT ON DILUTION! Where stock sulphuric acid is 18 Molar, add 13.9 ml to 86 ml H₂O. Store at room temperature.

9. Other:

Microplates, 96-well Immulon 4-HBX (<http://www.labsystems.fi>)
Microplate washer (optional)
Microplate reader.

Assay Procedure:

1. Coating of plates:

Dilute the capture antibody 1/100 in coating buffer (preferably in a polypropylene tube) and immediately add 100 µl to every well in the plate. Incubate for 2 hours @ 22°C or overnight @ 2-8°C.

2. Blocking:

Empty contents of plate and add 150 µl of blocking buffer to every well and incubate for 60 minutes @ 22°C.

Wash plate X 3 with wash buffer.

3. Samples:

Reference plasma is diluted 1/50 (100%) then serial 1/2's down to 1/1600 (3.13%). Sample plasmas are diluted 1/100, 1/200 & 1/400. All dilutions are made in HBS-BSA-T20 sample diluent. Apply 100 µl/well and incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

4. Detecting Antibody:

Dilute the detecting antibody 1/100 in HBS-BSA-T20 sample diluent and apply 100 µl to each well. Incubate plate @ 22°C for 60 minutes. Wash plate X 3 with wash buffer.

5. OPD Substrate:

Apply 100 µl of freshly prepared OPD substrate to every well. Allow colour to develop for 5-10 minutes then stop colour reaction with the addition of 50 µl/well of 2.5 M H₂SO₄. The plate can be read at a wavelength of 490 nm.

Calculation of Results:

The construction of a proper reference curve is of no less importance than any other aspect of the assay. A reference curve should be constructed by plotting the known concentration of standards versus absorbance. This can be done manually using graph paper, or by using curve-fitting computer software. In our experience, the dose response curves of most immunoassays tend to be sigmoid in shape. Although linear regions can be identified within the curve, the best overall fit is often obtained using an algorithm that provides a weighted theoretical model of fit throughout the entire curve, such as a 4-parameter or 5-parameter logistic curve fit^{4,5}. In general, the simplest model that defines the concentration-response relationship should be used⁶.

The "back-fit" test is a simple and reliable method to determine if a curve-fitting method is appropriate. In this test, the apparent concentrations for the absorbance values of each standard point are read from the reference curve. The derived values are compared to the assigned values. An appropriate curve fitting method will produce derived values that closely match assigned values throughout the range of the curve, within user-defined limits⁶. The coefficient of determination (R^2) is a valuable indicator of the overall fit, but should not be used by itself in the selection of a curve fitting method, as a poor fit in a particular region of the curve may not be evident from this value alone^{5,6}.

In the quality control of this product we have determined that under the conditions described above, a reference curve that is constructed using serial dilutions of normal pooled plasma, will produce a correlation coefficient (R^2) of at least 0.980 using a log-log fit, and an R^2 of at least 0.990 using a 4-parameter logistic curve fit algorithm. However, the performance characteristics of in-house assays developed using this product in other laboratories may vary slightly from ours. Different curve fitting methods may be employed but we recommend that the back-fit test be applied as evidence that the fitting method is appropriate.

Technical Notes:

- This paired antibody product is intended to facilitate the end user in establishing an in-house immunoassay for research purposes only. It must not be used for diagnostic applications. Assay validation is the responsibility of the end user and should be done according to user-defined protocols⁶.
- Reference calibrators should be of the same matrix and anticoagulant as the samples to be tested (example serum or plasma, citrate or EDTA).
- Do not use samples diluted less than 1/10, as falsely high readings may result.
- The optimal colour development time should be determined empirically as the time required to obtain an absorbance of at least 1.000 at 490 nm for the 100% reference point, not to exceed 20 minutes.
- Rheumatoid factor in samples may interfere in ELISA by binding to the capture and/or detecting antibodies.
- The wells should not be allowed to become dry. Keep plate covered or in a humid chamber during incubations.
- Antibodies are supplied in a 50% glycerol solution and can be centrifuged briefly in a micro-centrifuge to gather residual reagent from the cap and walls of the tube.

References:

1. Willemin WA, Minnema M, Meijers JCM, Roem D, Erenberg AJM, Nuijens JH, ten Cate H, Hack EC; Inactivation of Factor XIa in Human Plasma Assessed by Measuring Factor XIa-Protease Inhibitor Complexes: Major Role for C1-Inhibitor. *Blood* 85:1517, 1995.
2. DeLa Cadena R, Watchtfoegel YT, Colman RW, in Hemostasis and Thrombosis, 3rd Edition, eds. RW Colman, J Hirsh, VJ Marder and EW Salzman, pp. 219-240, J.B. Lippincott Co., Philadelphia, 1994.
3. Baglia FA, Seaman FS, Walsh, PN; The Apple 1 and 4 domains of Factor XI Act to Synergistically Promote the Surface-Mediated Activation of Factor XI by Factor XIIa. *Blood* 85:2078, 1995.
4. Nix,B, Wild D, in Immunoassays, A Practical Approach, editor J.P. Gosling, pp. 239-261, Oxford University Press, 2000.
5. NCCLS. Evaluation of the Linearity of Quantitative Analytical Methods; Proposed Guideline - Second Edition. NCCLS Document EP6-P2 (ISBN 1-56238-446-5, NCCLS, Wayne, Pennsylvania USA, 2001.
6. FDA Guidance for Industry. Bioanalytical Method Validation; May 2001, available on the internet: www.fda.gov/cder/guidance/index.htm

Related Products:

Cat #:	GAFXI-IG	Goat anti-human F.XI, whole IgG from antiserum
Cat #:	GAFXI-AP	Goat anti-human F.XI, affinity purified IgG
Cat #:	GAFXI-HRP	Goat anti-human F.XI, IgG-peroxidase conjugate
Cat #:	SAFXI-IG	Sheep anti-human F.XI, whole IgG from antiserum
Cat #:	SAFXI-AP	Sheep anti-human F.XI, affinity purified IgG
Cat #:	FXI -EIA	Paired Antibody set for ELISA of F.XI, 5 x 96 test
Cat #:	FXI-DP	Plasma deficient in F.XI, immune-depleted